



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

BYK - Gardner
c/o BYK-Gardner USA
9104 Guilford Road
Columbia, MD 21046, USA
(and satellite locations as listed on the scope)

Fulfills the requirements of

ISO/IEC 17025:2017

In the fields of

CALIBRATION and TESTING

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 28 April 2027
Certificate Number: AC-1534



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

BYK – Gardner

Dr. Ulrich Neuhäusler, 17025 Quality Manager
Ulrich.Neuhaeusler@altana.com
Phone: +49 (8171) 3493 357

This scope applies to the following locations:

Headquarter

c/o BYK – Gardner GmbH
Lausitzerstr.8, 82538 Geretsried, Germany

Headquarter USA

c/o BYK–Gardner USA
9104 Guilford Rd., Columbia, MD 21046, USA

BYK-Gardner Service Point Spain

c/o Actega Artística S.A
Calle Balmes 8, Suite: 3º 2ª, 08291
Ripollet, Spain

BYK–Gardner Service Point France

c/o ACTEGA Rhenacoat SAS
27 Allée du Chargement,
59650 Villeneuve d’Ascq, France

BYK-Gardner Service Point China

c/o BYK (Tongling) Co., Ltd. Shanghai Branch
Block 6A, Building A, No 88 Hong Cao Road,
Xuhui District, Shanghai 200233, P.R China

BYK-Gardner Service Point South Latin America

c/o MAST COMERCIAL E IMPORTADORA LTDA
Rua Itaporanga, 340-B, Bairro Paraíso,
Santo André – SP, 09190-640, Brazil

BYK-Gardner Service Point Japan

c/o TETSUTANI CO., LTD
Chuo-ku, Osaka, Tokui cho 2-2-2, Japan

BYK-Gardner Service Point Austria, Hungary, Slovenia

c/o FRIEDRICH W. BLOCH GmbH
Wagramerstrasse 201, 1210 Vienna, Austria

BYK-Gardner Service Point India

c/o IMCD India Private Limited
24th Floor, 29, The Ruby,
Senapati Bapat Marg, Dadar West
Mumbai, MH. Pin.: 400 028 India

BYK-Gardner Service Point North Latin America

c/o Exacolor Laboratories, S.A. de C.V.
Barranquilla No. 150, Monterrey,
Nuevo León, México, 64840

CALIBRATION AND TESTING

ISO/IEC 17025 Accreditation Granted: **24 April 2025**

Certificate Number: **AC-1534**

Certificate Expiry Date: **28 April 2027**

CALIBRATION

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Digital Film Thickness Units ^{1,3}	Up to 100 μm (>100 to 250) μm (>250 to 3 000) μm	2.9 μm 4.6 μm 11 μm	Comparison to Certified Shims	All
Digital Film Thickness micro-Tri-gloss ^{1,3}	Up to 100 μm (>100 to 250) μm (>250 to 3 000) μm	9.9 μm 11 μm 15 μm	Comparison to Certified Shims	All
Film Thickness Shims	Up to 100 μm (>100 to 250) μm (>250 to 3 000) μm	0.4 μm 0.6 μm 0.8 μm	Comparison to Micrometer	Columbia, MD Geretsried, Germany
Adhesion Tape Test Roller Length Weight Hardness, Shore A	(1 to 150) mm (1 to 3 000) g (70 to 90) Duro	0.1 mm 1.2 g 1.8 Duro	Comparison to Caliper Scale Durometer	Columbia, MD Geretsried, Germany

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Density Cups ³	(8 to 101) ml	0.1 % of reading	Indirect Verification using Scale, Thermometer	Columbia, MD Geretsried, Germany
Kinematic Viscosity ^{3,4} Ford Cups DIN Cups ISO Cups Zahn Cups	Drain Time (10 to 100) s	2.5 % of drain time 2 % of drain time 2 % of drain time 2 % of drain time	Comparison to Timer	Columbia, MD Geretsried, Germany

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Rotational Viscosity ^{3,7}	Up to 250 rpm (1 to 100) cP (101 to 1 000) cP (1 001 to 320 000) cP	2 cP 6 cP 40 cP	Indirect Verification using Thermometer, Certified Oil	Columbia, MD Geretsried, Germany Santo André, Brazil
	Up to 250 rpm (40 to 141) KU	3 KU		

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Gloss ^{1,2}	(0 to 10) GU (>10 to 94) GU (>94 to 100) GU (158-168 GU at 20°) (148-158 GU at 45°) (128-138 GU at 60°) (108-118 GU at 75°) (100-110 GU at 85°)	0.21 GU 0.61 GU 0.41 GU 0.34 GU 0.34 GU 0.33 GU 0.31 GU 0.36 GU	Comparison to Gloss Standards or Instrumentation	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.25 a*: 0.10 b*: 0.10 ΔE*(CIELab): 0.29 ΔE*(CIELCH): 0.29 ΔE*(CIE94): 0.21 L*: 0.30 a*: 0.10 b*: 0.10 ΔE*(CIELab): 0.29 ΔE*(CIELCH): 0.29 ΔE*(CIE94): 0.29	Comparison to White Standard or Instrumentation	All

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.21 a*: 0.14 b*: 0.14 $\Delta E^*(\text{CIELab})$: 0.29 $\Delta E^*(\text{CIELCH})$: 0.29 $\Delta E^*(\text{CIE94})$: 0.29 L*: 0.42 a*: 0.14 b*: 0.14 $\Delta E^*(\text{CIELab})$: 0.46 $\Delta E^*(\text{CIELCH})$: 0.46 $\Delta E^*(\text{CIE94})$: 0.46	Comparison to Pale Grey Standard (BCRA) or Instrumentation	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.21 a*: 0.14 b*: 0.18 $\Delta E^*(\text{CIELab})$: 0.31 $\Delta E^*(\text{CIELCH})$: 0.31 $\Delta E^*(\text{CIE94})$: 0.31 L*: 0.35 a*: 0.14 b*: 0.14 $\Delta E^*(\text{CIELab})$: 0.40 $\Delta E^*(\text{CIELCH})$: 0.40 $\Delta E^*(\text{CIE94})$: 0.40	Comparison to Mid/Diff Standard (BCRA) or Instrumentation	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.42 a*: 0.20 b*: 0.21 $\Delta E^*(\text{CIELab})$: 0.51 $\Delta E^*(\text{CIELCH})$: 0.51 $\Delta E^*(\text{CIE94})$: 0.51 L*: 0.71 a*: 0.14 b*: 0.20 $\Delta E^*(\text{CIELab})$: 0.75 $\Delta E^*(\text{CIELCH})$: 0.75 $\Delta E^*(\text{CIE94})$: 0.75	Comparison to Deep Grey Standard (BCRA) or Instrumentation	All

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.28 a*: 0.26 b*: 0.21 ΔE*(CIELab): 0.44 ΔE*(CIELCH): 0.44 ΔE*(CIE94): 0.43 L*: 0.42 a*: 0.28 b*: 0.21 ΔE*(CIELab): 0.55 ΔE*(CIELCH): 0.55 ΔE*(CIE94): 0.55	Comparison to Deep Pink Standard (BCRA) or Instrumentation	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.42 a*: 0.35 b*: 0.71 ΔE*(CIELab): 0.90 ΔE*(CIELCH): 0.90 ΔE*(CIE94): 0.88 L*: 0.57 a*: 0.42 b*: 1.3 ΔE*(CIELab): 1.5 ΔE*(CIELCH): 1.5 ΔE*(CIE94): 1.4	Comparison to Red Standard (BCRA) or Instrumentation	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.28 a*: 0.28 b*: 0.52 ΔE*(CIELab): 0.65 ΔE*(CIELCH): 0.65 ΔE*(CIE94): 0.65 L*: 0.42 a*: 0.28 b*: 0.86 ΔE*(CIELab): 1.0 ΔE*(CIELCH): 1.0 ΔE*(CIE94): 0.98	Comparison to Orange Standard (BCRA) or Instrumentation	All

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.35 a*: 0.21 b*: 0.44 $\Delta E^*(CIE_{Lab})$: 0.60 $\Delta E^*(CIE_{LCH})$: 0.60 $\Delta E^*(CIE_{94})$: 0.60 L*: 0.42 a*: 0.21 b*: 0.42 $\Delta E^*(CIE_{Lab})$: 0.63 $\Delta E^*(CIE_{LCH})$: 0.63 $\Delta E^*(CIE_{94})$: 0.63	Comparison to Bright Yellow Standard (BCRA) or Instrumentation	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.30 a*: 0.24 b*: 0.30 $\Delta E^*(CIE_{Lab})$: 0.49 $\Delta E^*(CIE_{LCH})$: 0.49 $\Delta E^*(CIE_{94})$: 0.49 L*: 0.35 a*: 0.21 b*: 0.28 $\Delta E^*(CIE_{Lab})$: 0.49 $\Delta E^*(CIE_{LCH})$: 0.49 $\Delta E^*(CIE_{94})$: 0.49	Comparison to Green/Diff Green Standard (BCRA) or Instrumentation	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.28 a*: 0.28 b*: 0.21 $\Delta E^*(CIE_{Lab})$: 0.45 $\Delta E^*(CIE_{LCH})$: 0.45 $\Delta E^*(CIE_{94})$: 0.45 L*: 0.42 a*: 0.28 b*: 0.28 $\Delta E^*(CIE_{Lab})$: 0.58 $\Delta E^*(CIE_{LCH})$: 0.58 $\Delta E^*(CIE_{94})$: 0.58	Comparison to Cyan Standard (BCRA) or Instrumentation	All

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 1.2 a*: 1.8 b*: 1.4 ΔE*(CIELab): 2.6 ΔE*(CIELCH): 2.6 ΔE*(CIE94): 2.5 L*: 2.1 a*: 3.2 b*: 2.5 ΔE*(CIELab): 4.5 ΔE*(CIELCH): 4.5 ΔE*(CIE94): 4.2	Comparison to Deep Blue Standard (BCRA) or Instrumentation	All
Clarity ^{1,2,6}	Up to 100 %	0.2 %	Comparison to Clarity Standards or Instrumentation ASTM D1003 Illuminant A, ASTM D1003 Illuminant C	All
Transmission ^{1,2,6}	Up to 50 % (>50 to 100) %	0.37 % 0.07 %	Comparison to Transmission Standards or Instrumentation ASTM D1003 Illuminant A, ASTM D1003 Illuminant C, ISO 13468	All
Haze ^{1,2,6}	(0.1 to 1) % (>1 to 10) % (>10 to 100) %	0.1 % 0.1 % 0.2 %	Comparison to Haze Standards or Instrumentation ASTM D1003 Illuminant A, ASTM D1003 Illuminant C, ISO 13468	All
Wavescan / DOI – Waviness Scale ^{1,2,6}	(0 to 99)	1.7	Comparison to Orange Peel Standard or Instrumentation	All

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Optical Radiation – Illuminance ^{1,3}	(50 to 200) fc (540 to 2 200) lx	11 fc 120 lx	Comparison to Spectroradiometer	Columbia, MD Geretsried, Germany Ripollet, Spain Villeneuve d'Ascq, France Vienna, Austria
Optical Radiation - Color Temperature ^{1,3}	(2 250 to 3 500) K (3 800 to 4 350) K (6 300 to 6 700) K	55 K 74 K 82 K	Comparison to Spectroradiometer	Columbia, MD Geretsried, Germany Ripollet, Spain Villeneuve d'Ascq, France Vienna, Austria

Thermodynamic

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment	Locations ⁵
Thermocouple Simulation ³	Type K (50 to 500) °C	0.9 °C	Millivolt Simulation of Temperature using Universal Calibrator	Columbia, MD Geretsried, Germany Vienna, Austria Shanghai, P.R China Monterrey, Mexico
Temperature ³	(15 to 50) °C	0.65 °C	Comparison to Climatic Chamber	Columbia, MD Geretsried, Germany
Relative Humidity ³	(30 to 75) %RH	1.8 %RH	Comparison to Climatic Chamber	Columbia, MD Geretsried, Germany
Dewpoint ³	(10 to 30) °C	Calculated from temperature and humidity	Calculated from temperature and humidity	Columbia, MD Geretsried, Germany

TESTING

Mechanical

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Digital Film Thickness Units ^{1,3}	Up to 100 μm (>100 to 250) μm (>250 to 3 000) μm	2.9 μm 4.6 μm 11 μm	Certified Shims BYK-Gardner working instructions Customer defined procedures	All
Digital Film Thickness micro-Tri-gloss ^{1,3}	Up to 100 μm (>100 to 250) μm (>250 to 3 000) μm	9.9 μm 11 μm 15 μm	Certified Shims BYK-Gardner working instructions Customer defined procedures	All
Film Thickness Shims	Up to 100 μm (>100 to 250) μm (>250 to 3 000) μm	0.4 μm 0.6 μm 0.8 μm	Micrometer BYK-Gardner working instructions Customer defined procedures	Columbia, MD Geretsried, Germany
Adhesion Tape Test Roller Length Weight Hardness, Shore A	(1 to 150) mm (1 to 3 000) gr (70 to 90) Duro	0.1 mm 1.2 g 1.8 Duro	Vernier Caliper Scale Durometer BYK-Gardner working instructions Customer defined procedures	Columbia, MD Geretsried, Germany
Density Cups ³	(8 to 101) ml	0.1 % of reading	Scale, Thermometer BYK-Gardner working instructions Customer defined procedures	Columbia, MD Geretsried, Germany
Kinematic Viscosity ^{3,4} Ford Cups DIN Cups ISO Cups Zahn Cups	Drain Time (10 to 100) s	2.5 % of drain time 2 % of drain time 2 % of drain time 2 % of drain time	Thermometer, Timer, Certified Oil BYK-Gardner working instructions Customer defined procedures	Columbia, MD Geretsried, Germany

Mechanical

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Rotational Viscosity ^{3,7}	Up to 250 rpm (1 to 100) cP (101 to 1 000) cP (1 001 to 320 000) cP	2 cP 6 cP 40 cP	Thermometer, Certified Oil BYK-Gardner working instructions Customer defined procedures	Columbia, MD Geretsried, Germany Santo André, Brazil
	Up to 250 rpm (40 to 141) KU	3 KU		

Optical

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Gloss ^{1,2}	(0 to 10) GU (>10 to 94) GU (>94 to 100) GU (158-168 GU at 20°) (148-158 GU at 45°) (128-138 GU at 60°) (108-118 GU at 75°) (100-110 GU at 85°)	0.21 GU 0.61 GU 0.41 GU 0.34 GU 0.34 GU 0.33 GU 0.31 GU 0.36 GU	Gloss Standards BYK-Gardner working instructions Customer defined procedures	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.25 a*: 0.10 b*: 0.10 ΔE*(CIELab): 0.29 ΔE*(CIELCH): 0.29 ΔE*(CIE94): 0.21 L*: 0.30 a*: 0.10 b*: 0.10 ΔE*(CIELab): 0.29 ΔE*(CIELCH): 0.29 ΔE*(CIE94): 0.29	White Standard BYK-Gardner working instructions Customer defined procedures	All

Optical

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.21 a*: 0.14 b*: 0.14 $\Delta E^*(CIE_{Lab})$: 0.29 $\Delta E^*(CIE_{LCH})$: 0.29 $\Delta E^*(CIE_{94})$: 0.29 L*: 0.42 a*: 0.14 b*: 0.14 $\Delta E^*(CIE_{Lab})$: 0.46 $\Delta E^*(CIE_{LCH})$: 0.46 $\Delta E^*(CIE_{94})$: 0.46	Pale Grey Standard (BCRA) BYK-Gardner working instructions Customer defined procedures	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.21 a*: 0.14 b*: 0.18 $\Delta E^*(CIE_{Lab})$: 0.31 $\Delta E^*(CIE_{LCH})$: 0.31 $\Delta E^*(CIE_{94})$: 0.31 L*: 0.35 a*: 0.14 b*: 0.14 $\Delta E^*(CIE_{Lab})$: 0.40 $\Delta E^*(CIE_{LCH})$: 0.40 $\Delta E^*(CIE_{94})$: 0.40	Mid/Diff Standard (BCRA) BYK-Gardner working instructions Customer defined procedures	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.42 a*: 0.20 b*: 0.21 $\Delta E^*(CIE_{Lab})$: 0.51 $\Delta E^*(CIE_{LCH})$: 0.51 $\Delta E^*(CIE_{94})$: 0.51 L*: 0.71 a*: 0.14 b*: 0.20 $\Delta E^*(CIE_{Lab})$: 0.75 $\Delta E^*(CIE_{LCH})$: 0.75 $\Delta E^*(CIE_{94})$: 0.75	Deep Grey Standard (BCRA) BYK-Gardner working instructions Customer defined procedures	All

Optical

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.28 a*: 0.26 b*: 0.21 ΔE*(CIELab): 0.44 ΔE*(CIELCH): 0.44 ΔE*(CIE94): 0.43 L*: 0.42 a*: 0.28 b*: 0.21 ΔE*(CIELab): 0.55 ΔE*(CIELCH): 0.55 ΔE*(CIE94): 0.55	Deep Pink Standard (BCRA) BYK-Gardner working instructions Customer defined procedures	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.42 a*: 0.35 b*: 0.71 ΔE*(CIELab): 0.90 ΔE*(CIELCH): 0.90 ΔE*(CIE94): 0.88 L*: 0.57 a*: 0.42 b*: 1.3 ΔE*(CIELab): 1.5 ΔE*(CIELCH): 1.5 ΔE*(CIE94): 1.4	Red Standard (BCRA) BYK-Gardner working instructions Customer defined procedures	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.28 a*: 0.28 b*: 0.52 ΔE*(CIELab): 0.65 ΔE*(CIELCH): 0.65 ΔE*(CIE94): 0.65 L*: 0.42 a*: 0.28 b*: 0.86 ΔE*(CIELab): 1.00 ΔE*(CIELCH): 1.00 ΔE*(CIE94): 0.98	Orange Standard (BCRA) BYK-Gardner working instructions Customer defined procedures	All

Optical

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.35 a*: 0.21 b*: 0.44 $\Delta E^*(CIE_{Lab})$: 0.60 $\Delta E^*(CIE_{LCH})$: 0.60 $\Delta E^*(CIE_{94})$: 0.60 L*: 0.42 a*: 0.21 b*: 0.42 $\Delta E^*(CIE_{Lab})$: 0.63 $\Delta E^*(CIE_{LCH})$: 0.63 $\Delta E^*(CIE_{94})$: 0.63	Bright Yellow Standard (BCRA) BYK-Gardner working instructions Customer defined procedures	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.30 a*: 0.24 b*: 0.30 $\Delta E^*(CIE_{Lab})$: 0.49 $\Delta E^*(CIE_{LCH})$: 0.49 $\Delta E^*(CIE_{94})$: 0.49 L*: 0.35 a*: 0.21 b*: 0.28 $\Delta E^*(CIE_{Lab})$: 0.49 $\Delta E^*(CIE_{LCH})$: 0.49 $\Delta E^*(CIE_{94})$: 0.49	Green/Diff Green Standard (BCRA) BYK-Gardner working instructions Customer defined procedures	All
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 0.28 a*: 0.28 b*: 0.21 $\Delta E^*(CIE_{Lab})$: 0.45 $\Delta E^*(CIE_{LCH})$: 0.45 $\Delta E^*(CIE_{94})$: 0.45 L*: 0.42 a*: 0.28 b*: 0.28 $\Delta E^*(CIE_{Lab})$: 0.58 $\Delta E^*(CIE_{LCH})$: 0.58 $\Delta E^*(CIE_{94})$: 0.58	Cyan Standard (BCRA) BYK-Gardner working instructions Customer defined procedures	All

Optical

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Color / Spectrophotometer ^{1,2}	8°:di / 8°:de: (380 to 780 nm) 0°:45°a (380 to 780 nm)	L*: 1.2 a*: 1.8 b*: 1.4 ΔE*(CIELab): 2.6 ΔE*(CIELCH): 2.6 ΔE*(CIE94): 2.5 L*: 2.1 a*: 3.2 b*: 2.5 ΔE*(CIELab): 4.5 ΔE*(CIELCH): 4.5 ΔE*(CIE94): 4.2	Deep Blue Standard (BCRA) BYK-Gardner working instructions Customer defined procedures	All
Clarity ^{1,2,6}	Up to 100 %	0.2 %	Clarity Standards ASTM D1003 Illuminant A, ASTM D1003 Illuminant C, BYK-Gardner working instructions Customer defined procedures	All
Transmission ^{1,2,6}	Up to 50 % (>50 to 100) %	0.37 % 0.07 %	Transmission Standards ASTM D1003 Illuminant A, ASTM D1003 Illuminant C, ISO 13468, BYK-Gardner working instructions Customer defined procedures	All

Optical

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	Locations ⁵
Haze ^{1,2,6}	(0.1 to 1) % (>1 to 10) % (>10 to 100) %	0.1 % 0.1 % 0.2 %	Haze Standards ASTM D1003 Illuminant A, ASTM D1003 Illuminant C, ISO 13468, BYK-Gardner working instructions Customer defined procedures	All
Wavescan / DOI – Waviness Scale ^{1,2,6}	(0 to 99)	1.7	Orange Peel Standards BYK-Gardner working instructions Customer defined procedures	All
Optical Radiation – Illuminance ^{1,3}	(50 to 200) fc (540 to 2 200) lx	11 fc 120 lx	Spectroradiometer BYK-Gardner working instructions Customer defined procedures	Columbia, MD Geretsried, Germany Ripollet, Spain Villeneuve d'Ascq, France Vienna, Austria
Optical Radiation - Color Temperature ^{1,3}	(2 250 to 3 500) K (3 800 to 4 350) K (6 300 to 6 700) K	55 K 74 K 82 K	Spectroradiometer BYK-Gardner working instructions Customer defined procedures	Columbia, MD Geretsried, Germany Ripollet, Spain Villeneuve d'Ascq, France Vienna, Austria

Thermal

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment	Locations ⁵
Thermocouple Simulation ³	Type K (50 to 500) °C	0.9 °C	Universal Calibrator BYK-Gardner working instructions Customer defined procedures	Columbia, MD Geretsried, Germany Vienna, Austria Shanghai, P.R China Monterrey, Mexico

Thermal

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment	Locations ⁵
Temperature	(15 to 50) °C	0.65 °C	Climatic Chamber BYK-Gardner working instructions Customer defined procedures	Columbia, MD Geretsried, Germany
Relative Humidity	(30 to 75) %RH	1.8 %RH	Climatic Chamber BYK-Gardner working instructions Customer defined procedures	Columbia, MD Geretsried, Germany
Dewpoint	(10 to 30) °C	Calculated from temperature and humidity	Calculated from temperature and humidity BYK-Gardner working instructions Customer defined procedures	Columbia, MD Geretsried, Germany

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration/testing services are available for this parameter; based on strict protocols, the same uncertainties are achieved on-site.
2. Applies to both instruments and standards.
3. Applies to instruments only.
4. Drain time of certified calibration oil.
5. The capabilities of all sites are identical using the same procedures and equipment, under the same environment conditions.
6. Unit-less measure expressed as a percentage (not a percentage of reading).
7. Ranges of rotational viscometers above 125 000 cP and above 105 KU are extrapolated mathematically, KU is Krebs-Stromer Unit.



Jason Stine, Vice President